



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/599,169	09/15/2008	Shinya Nagata	5553NA4-1	5276
62574	7590	12/30/2011	EXAMINER	
Jason H. Vick Sheridan Ross, PC Suite # 1200 1560 Broadway Denver, CO 80202			BEHRINGER, LUTHER G	
			ART UNIT	PAPER NUMBER
			3766	
			NOTIFICATION DATE	DELIVERY MODE
			12/30/2011	ELECTRONIC

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

jvick@sheridanross.com

Office Action Summary	Application No. 10/599,169	Applicant(s) NAGATA ET AL.	
	Examiner LUTHER BEHRINGER	Art Unit 3766	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 1 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 21 September 2006.
- 2a) ☐ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ An election was made by the applicant in response to a restriction requirement set forth during the interview on ____; the restriction requirement and election have been incorporated into this action.
- 4) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 5) ☒ Claim(s) 22-42 is/are pending in the application.
- 5a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 6) ☐ Claim(s) ____ is/are allowed.
- 7) ☐ Claim(s) ____ is/are rejected.
- 8) ☐ Claim(s) ____ is/are objected to.
- 9) ☒ Claim(s) 22-42 are subject to restriction and/or election requirement.

Application Papers

- 10) ☐ The specification is objected to by the Examiner.
- 11) ☐ The drawing(s) filed on ____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 12) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 13) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. ____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. ____. |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date ____. | 6) <input type="checkbox"/> Other: ____. |

DETAILED ACTION

Election/Restrictions

1. Restriction is required under 35 U.S.C. 121 and 372.

This application contains the following inventions or groups of inventions which are not so linked as to form a single general inventive concept under PCT Rule 13.1.

In accordance with 37 CFR 1.499, applicant is required, in reply to this action, to elect a single invention to which the claims must be restricted.

Group I, claim(s) 22 – 27 and 31 – 35, drawn to a garment for measuring biological information.

Group II, claim(s) 28 – 30, 37 and 41, drawn to a cardiogram analysis device.

Group II, claim(s) 36, 38 – 40 and 42, drawn to a respiratory information analysis device.

2. The groups of inventions listed above do not relate to a single general inventive concept under PCT Rule 13.1 because, under PCT Rule 13.2, they lack the same or corresponding special technical features for the following reasons:

The inventions of groups I, II and III lack unity of invention because even though the inventions of these groups require the technical feature of a garment having electrodes, this technical feature is not a special technical feature as it does not make a contribution over the prior art in view of US 276,078, US 4,580,572, US 6,047,203, US 5,353,793, and US 5,178,151. The incorporation of electrodes into garments is old and well known, as is demonstrated by the references cited performing similarly to the

Art Unit: 3766

claimed invention as far back as 1986. Indeed, electrodes have been incorporated into belts at least since 1883 as is demonstrated by US 276,078. Therefore, incorporation of electrodes into garments lacks novelty and does not make a contribution to the art.

Should applicant elect **group I**, a further species election is deemed necessary.

3. REQUIREMENT FOR UNITY OF INVENTION

As provided in 37 CFR 1.475(a), a national stage application shall relate to one invention only or to a group of inventions so linked as to form a single general inventive concept ("requirement of unity of invention"). Where a group of inventions is claimed in a national stage application, the requirement of unity of invention shall be fulfilled only when there is a technical relationship among those inventions involving one or more of the same or corresponding special technical features. The expression "special technical features" shall mean those technical features that define a contribution which each of the claimed inventions, considered as a whole, makes over the prior art.

The determination whether a group of inventions is so linked as to form a single general inventive concept shall be made without regard to whether the inventions are claimed in separate claims or as alternatives within a single claim. See 37 CFR 1.475(e).

WHEN CLAIMS ARE DIRECTED TO MULTIPLE CATEGORIES OF INVENTIONS

As provided in 37 CFR 1.475(b), a national stage application containing claims to different categories of invention will be considered to have unity of invention if the claims are drawn only to one of the following combinations of categories:

- (1) A product and a process specially adapted for the manufacture of said product; or
- (2) A product and process of use of said product; or
- (3) A product, a process specially adapted for the manufacture of the said product, and a use of the said product; or
- (4) A process and an apparatus or means specifically designed for carrying out the said process; or
- (5) A product, a process specially adapted for the manufacture of the said product, and an apparatus or means specifically designed for carrying out the said process.

Otherwise, unity of invention might not be present. See 37 CFR 1.475(c).

4. This application contains claims directed to more than one species of the generic invention. These species are deemed to lack unity of invention because they are not so linked as to form a single general inventive concept under PCT Rule 13.1.

The species are as follows:

A garment having electrodes for cardiogram analysis (see claim 27)

A garment having electrodes for respiratory information analysis (see claim 35)

Applicant is required, in reply to this action, to elect a single species to which the claims shall be restricted if no generic claim is finally held to be allowable. The reply must also identify the claims readable on the elected species, including any claims subsequently added. An argument that a claim is allowable or that all claims are generic is considered non-responsive unless accompanied by an election.

Upon the allowance of a generic claim, applicant will be entitled to consideration of claims to additional species which are written in dependent form or otherwise require all the limitations of an allowed generic claim. Currently, the following claim(s) are generic: Claim 22

Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the requirement may be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

The election of an invention or species may be made with or without traverse. To preserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse. Traversal must be presented at the time of election in order to be considered timely. Failure to timely traverse the requirement will result in the loss of right to petition under 37 CFR 1.144. If claims are added after the election, applicant must indicate which of these claims are readable on the elected invention or species.

Should applicant traverse on the ground that the inventions have unity of invention (37 CFR 1.475(a)), applicant must provide reasons in support thereof. Applicant may submit evidence or identify such evidence now of record showing the inventions to be obvious variants or clearly admit on the record that this is the case. Where such evidence or admission is provided by applicant, if the examiner finds one of the inventions unpatentable over the prior art, the evidence or admission may be used in a rejection under 35 U.S.C. 103(a) of the other invention.

5. Applicant is reminded that upon the cancellation of claims to a non-elected invention, the inventorship must be amended in compliance with 37 CFR 1.48(b) if one or more of the currently named inventors is no longer an inventor of at least one claim remaining in the application. Any amendment of inventorship must be accompanied by a request under 37 CFR 1.48(b) and by the fee required under 37 CFR 1.17(i).

Any inquiry concerning this communication or earlier communications from the examiner should be directed to LUTHER BEHRINGER whose telephone number is (571)270-3868. The examiner can normally be reached on Mon - Thurs 9:00 - 6:30; 2nd Friday 9:00 - 5:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Carl Layno can be reached on (571) 272-4949. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3766

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/CARL H LAYNO/
Supervisory Patent Examiner, Art Unit 3766

/Luther G Behringer/
Examiner, Art Unit 3766